



BD Advantage™ 2 PCR Enzyme System User Manual

PT3281-1 (PR33798)
Published 11 April 2003



Table of Contents

I.	Introduction	3
II.	List of Components	6
III.	Additional Materials Required	7
IV.	General Considerations	9
	A. Primer Design	9
	B. Template Quality	9
	C. Amplicon Size	9
	D. Background Amplification	10
	E. Good PCR Practices	10
	F. Additional Considerations	11
V.	BD Advantage™ 2 PCR Procedure	12
	A. Control PCR Reactions	12
	B. Recommended Cycling Parameters for all BD Advantage 2 PCR Products	13
	C. Recommendations for Electrophoresis	14
	D. Using 10X BD Advantage 2 SA PCR Buffer	14
VI.	Troubleshooting Guide	15
VII.	References	19
VIII.	Related Products	19

I. Introduction

The **BD Advantage™ 2 Polymerase Mix** and **BD Advantage™ 2 PCR Kit** (which includes the Polymerase Mix) produce efficient, accurate, and convenient amplification of DNA from any template.

The **BD Advantage 2 Polymerase Mix** is comprised of BD TITANIUM™ *Taq* DNA Polymerase—a nuclease-deficient N-terminal deletion of *Taq* DNA polymerase plus BD TaqStart™ Antibody to provide automatic hot-start PCR (Kellogg *et al.*, 1994)—and a minor amount of a proofreading polymerase.

BD TITANIUM *Taq* provides the most sensitive and robust capabilities of any *Taq*-derived polymerase. Its increased sensitivity and robust nature are especially useful for amplifying a wide size range of DNA fragments, cDNAs of rare transcripts, or products from complex templates.

The higher yields and increased sensitivity that BD TITANIUM *Taq* provides translate into two major advantages over conventional polymerases. First, targets can be amplified using fewer PCR cycles, saving time and lowering background in any given experiment. Second, in situations where the amplification target is present at extremely low levels (e.g., amplifying a rare cDNA in an RT-PCR experiment or detecting viral nucleic acid), the high sensitivity obtained with BD TITANIUM *Taq* allows successful amplification of your target where other polymerases fail.

BD TITANIUM *Taq* allows you to perform PCR without tedious buffer optimization. In any given reaction, BD TITANIUM *Taq* tolerates a wide range of Mg²⁺ concentrations resulting in higher yields, especially for longer amplicons. Magnesium is already included at a set concentration in the **BD Advantage 2 PCR Buffer**, eliminating the need to add Mg²⁺ as a separate component during reaction setup. In contrast, native *Taq* polymerase only functions well over a narrow [Mg²⁺] range, and different reactions may require different concentrations of Mg²⁺. By eliminating the need to perform experiments for determining the optimal Mg²⁺ concentration, BD TITANIUM *Taq* saves considerable time and effort. In cases where background amplification is a problem, we recommend the use of 10X BD Advantage™ 2 SA PCR Buffer.

The **10X BD Advantage™ 2 SA PCR Buffer**, available separately (#639147, #639148), has been optimized for applications using genomic DNA, and in particular with products less than 1-2 kb. We have found yields to be comparable to that obtained with BD Advantage 2 PCR Buffer for relatively small amplicons; however, for amplicons > 2 kb, we recommend the use of the BD Advantage 2 PCR Buffer for best results. The 10X BD Advantage 2 SA PCR Buffer contains Mg²⁺ at a concentration optimized for most applications, particularly those utilizing a genomic DNA template. In addition, the Advantage 2 SA PCR Buffer is recommended when background amplification is present, which may sometimes result as a by-product of the sensitivity and robustness of the BD Advantage 2 Polymerase Mix.

I. Introduction *continued*

The BD Advantage™ Polymerase Systems

The simultaneous use of two different DNA polymerases (primary and proofreading) in a PCR reaction allows amplification of significantly longer fragments in a process known as long and accurate PCR (or long-distance PCR [LD PCR]; Barnes, 1994; Cheng *et al.*, 1994). However, the usefulness of two-enzyme systems is not limited to LD PCR. In fact, the efficiency of *most* PCR reactions can be significantly improved by using the two-enzyme combination.

The BD Advantage 2 system offers three primary benefits over conventional, single-polymerase PCR:

- **Increased range.** Whereas the upper limit of conventional PCR using a *Taq* polymerase is ~3 kb (and much lower in many applications), BD Advantage 2 gives consistent and efficient amplifications of up to 18 kb or more when using two nondegenerate primers of sufficient length to amplify an abundant, noncomplex template. It can also amplify high-complexity (i.e., genomic) DNA templates up to 6 kb. The absolute upper limit in any particular application will depend on the particular primers, the template used, and other experimental variables.
- **Increased fidelity.** The inclusion of a minor amount of a proofreading polymerase results in an error rate that is 3-fold lower than that of conventional PCR using *Taq* alone (Barnes, 1994; Frey *et al.*, 1995; Nelson *et al.*, 1995). In our studies, BD Advantage 2 exhibits an error rate of 25 errors per 100,000 bp after 25 PCR cycles. Note that the presence of organic solvents or salts in the reaction can decrease fidelity. High fidelity is a particularly important feature when the amplification products will be used in subsequent experiments (e.g., cloning, sequencing, functional assays, expression systems, etc.).
- **Increased efficiency and greater yields.** While range and fidelity are the most commonly noted aspects of long and accurate PCR, the use of a two-polymerase system also increases the efficiency and yield—and therefore the sensitivity—of *all* PCR assays, even for templates that are well within the range of conventional PCR. While DNA polymerases with proofreading activity offer better accuracy than LD PCR methods when used alone, they lack the increased efficiency and size-range flexibility possible with BD Advantage 2 system for long and accurate PCR.

Automatic hot start with BD TaqStart™ Antibody

The BD Advantage 2 system contains built-in, hot-start PCR from BD TaqStart Antibody included in the polymerase mix. Antibody-mediated hot start using BD TaqStart Antibody has been shown to significantly improve the efficiency and specificity of PCR amplifications by reducing background DNA synthesis (Kellogg *et al.*, *Clontechniques*, 1994; Kellogg *et al.*, *BioTechniques*, 1994). Specifically, this antibody reduces or eliminates nonspecific amplification products and primer-dimer artifacts created prior to the onset of thermal cycling.

I. Introduction *continued*

BD TaqStart is a neutralizing monoclonal antibody that recognizes both native *Taq* and N-terminal deletions such as BD TITANIUM *Taq*. The antibody inhibits enzymatic activity during PCR reaction setup at ambient temperatures. Polymerase activity is restored at the onset of thermal cycling because the antibody is denatured at high temperatures. The loss of inhibition is complete and irreversible, so the polymerase regains its full enzymatic activity for PCR.

Besides increased specificity and sensitivity, the built-in hot start in the BD Advantage 2 Polymerase Mix offers convenience. Other methods of hot start require extra steps such as the addition and premelting of wax beads or the addition of a critical component after the initial denaturation. These extra steps are inconvenient and introduce a potential source of cross-contamination. In contrast, BD TaqStart provides automatic hot-start PCR with virtually no risk of cross-contamination. Thus, BD TaqStart Antibody provides all the advantages of hot-start PCR with none of the disadvantages of other hot-start methods. The antibody comes already included in the BD Advantage 2 Polymerase Mix; there is no need to add it as a separate reagent during PCR setup.

Recommended uses for BD Advantage™ 2 Products

The BD Advantage 2 Mix and Kit are the recommended polymerase systems for use in applications involving RACE, RT-PCR, cDNA synthesis and library construction, cDNA subtraction and differential display, high-performance cloning, and RNA fingerprinting. The BD Advantage 2 System has been optimized for use with all BD Biosciences Clontech PCR-based application kits, including BD SMART™ cDNA Library Construction, BD SMART™ PCR cDNA Synthesis, BD SMART™ RACE cDNA Amplification, BD Clontech PCR-Select™ Subtraction Kits, BD Marathon™ cDNA Amplification, and BD Delta™ Differential Display. For genomic applications, including BD GenomeWalker™ Kits, we recommend BD Advantage™ Genomic Kits. See Related Products (Section VIII) for ordering information.

II. List of Components

BD Advantage™ 2 PCR Kit (Cat. No. 639207 & 639206)

Store all components at -20°C . Enough reagents are supplied for 30 or 100 PCR reactions of 50 μl each.

#639207 #639206

30 rxns 100 rxns

- **30 μl 100 μl 50X BD Advantage™ 2 Polymerase Mix**

Includes BDTITANIUM *Taq* DNA Polymerase, a small amount of proofreading polymerase, and BD TaqStart Antibody (1.1 $\mu\text{g}/\mu\text{l}$) in the following storage buffer:

Concentration in 50X mix	Component	Final rxn concentration
50 %	Glycerol	1.0 %
15 mM	Tris-HCl (pH 8.0)	0.3 mM
75 mM	KCl	1.5 mM
0.05 mM	EDTA	1.0 μM

- **200 μl 600 μl 10X BD Advantage™ 2 PCR Buffer**

Concentration in 10X mix	Component	Final rxn concentration
400 mM	Tricine-KOH (pH 8.7 at 25°C)	40 mM
150 mM	KOAc	15 mM
35 mM	$\text{Mg}(\text{OAc})_2$	3.5 mM
37.5 $\mu\text{g}/\text{ml}$	BSA	3.75 $\mu\text{g}/\text{ml}$
0.05 %	Tween 20	0.005 %
0.05 %	Nonidet-P40	0.005 %

- **200 μl 600 μl 10X BD Advantage™ 2 SA PCR Buffer**

- **50 μl 120 μl 50X dNTP Mix** (10 mM each of dATP, dCTP, dGTP, and dTTP; final rxn concentration: 0.2 mM each)

- **30 μl 100 μl Control DNA Template** (100 ng/ μl)
Calf Thymus DNA

- **30 μl 100 μl Control Primer Mix** (10 μM each)

5' primer 5'-GCAACTGCAGGAAGAGCAAGAAATGCA-3'

3' primer 5'-TGGCACGGCCATAAGAGGTAGATGTCA-3'

- **2.5 ml 5.0 ml PCR-Grade Water**

II. List of Components *continued*

BD Advantage™ 2 Polymerase Mix (Cat. No. 639201 & 639202)

Store all components at -20°C . Enough reagents are supplied for 100 or 500 PCR reactions of 50 μl each.

- | | | |
|-----------------|-----------------|--|
| #639201 | #639202 | |
| 100 rxns | 500 rxns | |
- **100 μl** **5 x 100 μl** **50X Advantage 2 Polymerase Mix**
See page 6 for component concentrations.
 - **600 μl** **5 x 600 μl** **10X Advantage 2 PCR Buffer**
See page 6 for component concentrations.
 - **600 μl** **5 x 600 μl** **10X Advantage 2 SA PCR Buffer**

10X BD Advantage™ 2 SA PCR Buffer (Cat. No. 639147 & 639148)

Store at -20°C . Enough buffer is supplied for 240 or 2,000 PCR reactions of 50 μl each.

- | | | |
|-----------------|-------------------|--|
| #639147 | #639148 | |
| 240 rxns | 2,000 rxns | |
- **2 x 600 μl** **10 ml** **10X BD Advantage™ 2 SA PCR Buffer**
Buffer alone.
- | Concentration
in 10X mix | Component | Final rxn
concentration |
|-----------------------------|-------------------|----------------------------|
| 100 mM | Tris-HCl (pH 8.5) | 10 mM |
| 500 mM | KCl | 50 mM |
| 20 mM | MgCl ₂ | 2 mM |

III. Additional Materials Required

The following reagents are not supplied.

- **Thermal cycler** (Perkin-Elmer GeneAmp 480 or equivalent)
- **Pipettors** dedicated for PCR
- **PCR pipette tips** suitable for use with the above pipettors and preferably equipped with hydrophobic filters.
- **DNA size markers** (See Section V.D.)
- **5X Stop/loading buffer** (Sambrook *et al.* [1989] provides several recipes)
- **PCR-grade water**
Avoid using autoclaved H₂O because recycled steam in some autoclaves can introduce contaminants that may interfere with PCR.

III. Additional Materials Required *continued*

- **0.5-ml PCR reaction tubes** (We recommend Perkin-Elmer GeneAmp 0.5-ml reaction tubes, Cat. No. N801-0737 or N801-0180.)
- [Optional] **Mineral oil** (We recommend Sigma Cat. No. M3516.)

If you have purchased the **BD Advantage 2 Polymerase Mix** alone (Cat. No. 639201, 639202), you will need the following:

- dNTP mix
- appropriate control template and primers

If you have purchased the **BD Advantage 2 SA PCR Buffer** alone (Cat. No. 639147, 639148), you will need the following:

- heat-stable DNA Polymerase
- dNTP mix
- appropriate control template and primers

IV. General Considerations

A. Primer Design

Primer design is the single largest variable in PCR applications and the single most important factor in determining the success or failure of PCR reactions. *Always check and recheck your primer design before constructing or ordering primers.* Visit alces.med.umn.edu/VGC.html on the web for helpful guidelines on primer design.

Length and G-C content: The BD Advantage 2 system can be used in a wide variety of PCR applications, and the constraints on primer design will vary from one application to the next. In general, however, primers should have a T_m of approximately 70°C to achieve optimal results in a two-step cycling program with a 68°C annealing/extension step. Therefore, whenever possible, primers should be *at least* 22 nucleotides (nt) long (25–30-mers are preferred) and should have a G-C content of 45–60%. Furthermore, the 3'-terminal ends of each primer should not be complementary to each other and should contain a low G-C content.

B. Template Quality

Because PCR amplification proceeds exponentially, many conventional PCR applications work well with templates of average or even low quality. In many applications (such as screening cDNA inserts with LD Insert Screening Amplimers), long and accurate PCR with BD Advantage 2 will also tolerate a wide range of template quality.

However, the longer or more complex the target, the more important template quality becomes. This is because the number of unnicked, full-length targets decreases as the target length increases, so poor-quality DNA will have very few large, unnicked targets. Furthermore, some depurination occurs when DNA is denatured during thermal cycling, and this can lead to truncated products. Therefore, it is particularly important to prepare high-quality, high molecular-weight DNA when amplifying large targets.

Template quality is also important when the highest possible sensitivity is needed. In cDNA applications such as RACE and RT-PCR protocols, incomplete reverse transcription can lead to an absence of product, truncated products, or a mix of truncated and full-length product, resulting in a smeared band on a gel. This problem can be minimized by ensuring that your starting material is of the highest quality. For 5' and 3' RACE and general PCR from cDNA, you can ensure the quality of your cDNA by using BD™ Marathon-Ready cDNA from BD Biosciences Clontech.

C. Amplicon Size

Yields are unaffected when amplifying products less than 2 kb, whether the 10X BD Advantage 2 SA PCR Buffer or the BD Advantage 2 PCR Buffer is used. We recommend initial use of the 10X BD Advantage 2 PCR Buffer; however, if a problem arises, in particular with amplicons less than 2 kb in length, then we recommend switching to the 10X BD Advantage 2 SA

IV. General Considerations *continued*

PCR Buffer. For products greater than 2 kb, we recommend the original BD Advantage 2 PCR Buffer be used (rather than the SA PCR Buffer).

D. Background Amplification

Due to the sensitivity and robustness of the BD Advantage 2 Polymerase Mix, non-specific background amplification may sometimes result. The 10X BD Advantage 2 Short Amplicon (SA) PCR Buffer has been specifically engineered to address this issue. We recommend the use of this buffer in those applications where some background amplification may result.

E. Good PCR Practices

1. Prepare reactions with dedicated pipettors in a dedicated work space

Due to the tremendous amplification power of PCR, minute amounts of contaminating DNA can produce nonspecific amplification; in some instances, contaminants can cause DNA bands even in the absence of added template DNA. We recommend that you use small aliquots of starting material to avoid contaminating your stocks. When performing PCR, you should wear gloves and set up your reactions in a dedicated lab area or noncirculating containment hood using dedicated pipettors, PCR pipette tips with hydrophobic filters, and dedicated solutions. We also recommend setting up a negative control reaction that does not contain any template. Finally, perform post-PCR analysis in a separate area using a separate set of pipettors.

2. Pipetting

Because of the small volumes used in PCR experiments and the potential for tube-to-tube variation, careful pipetting technique is extremely important. Always be sure that no extra solution is on the outside of the pipette tip before transfer. When adding solution to a tube, immerse the tip into the reaction mixture, deliver the contents from the pipette tip into the mixture, and pipet up and down several times.

3. Use a Master Mix

Assembling a Master Mix, which contains the appropriate volumes of all reagents required for multiple PCR reactions, saves time and greatly reduces tube-to-tube variation. If multiple templates are being tested with the same primers, include the primers in the Master Mix. If one template is being tested with multiple primer sets, include the template in the Master Mix. If you are setting up several sets of parallel samples, assemble multiple Master Mixes (e.g., each with a different set of primers). *The Master Mix should be thoroughly mixed before use (i.e., vortexed without bubbling).*

4. Always include positive and negative controls (i.e., H₂O instead of DNA template).

IV. General Considerations *continued*

F. Additional Considerations

1. Touchdown PCR

“Touchdown” PCR can significantly improve the specificity of many PCR reactions in a wide variety of applications (Don *et al.*, 1991; Roux, 1995). Briefly, touchdown PCR involves using an annealing/extension temperature that is several degrees (typically 3–10°C) higher than the T_m of the primers during the initial PCR cycles (typically 5–10 cycles). The annealing/extension temperature is then reduced to the primer T_m for the remaining PCR cycles.

2. BD TaqStart Antibody provides automatic hot start PCR

The use of a manual hot start or wax bead-based hot start is not required when using the BD Advantage 2 system. As discussed in the Introduction, hot start is automatic because the enzyme mix already contains BD TaqStart Antibody.

3. Half-life

The half-life of BD TITANIUM *Taq* depends on the specific reaction conditions used, but generally ranges from 20–40 min at 95°C.

4. Use of additives

BD TaqStart Antibody binds BD TITANIUM *Taq* DNA Polymerase with high affinity under the conditions described in this protocol. The addition of 2–5% DMSO will not interfere with TaqStart antibody function and may improve results in some instances. However, the addition of formamide or other cosolvents may disrupt TaqStart antibody function. Furthermore, excessive glycerol, solutes (e.g., salts), pH extremes, or other deviations from the recommended reaction conditions may reduce the effectiveness of the antibody and/or DNA polymerases.

5. BD Advantage 2 Polymerase Mix is not intended for certain applications

Because of the improved fidelity from long and accurate PCR, BD Advantage 2 is not recommended for mutagenesis protocols involving so-called “sloppy” PCR.

6. T/A Cloning

BD TITANIUM *Taq* PCR products are compatible with T/A cloning methods. However, for optimal ligation efficiencies, we recommend that you use PCR products immediately (<1 day) after amplification. The single 3' A-overhangs on the PCR products will degrade over time, reducing the efficiency. For best results, please observe the T/A cloning tips in Section V.C.

V. BD Advantage™ 2 PCR Procedure

PLEASE READ THE ENTIRE PROTOCOL BEFORE STARTING.

A. Control PCR Reactions

The following control PCR reactions should be performed in parallel with your experiments to ensure that the BD Advantage 2 Polymerase Mix is working properly. A positive control template and primers are provided in the BD Advantage 2 PCR Kit. When using the BD Advantage 2 Polymerase Mix with our other PCR-based applications kits, use the positive controls provided with those kits.

1. Place all components on ice and allow to thaw completely. Mix each component thoroughly before use.
2. Combine the following reagents in a 0.5-ml PCR tube.

Positive Control	Negative Control	
40 μ l	41 μ l	PCR-Grade Water
5 μ l	5 μ l	10X BD Advantage 2 PCR Buffer
1 μ l	---	Control DNA Template (100 ng/ μ l)
2 μ l	2 μ l	Control Primer Mix (10 μ M ea.)
1 μ l	1 μ l	50X dNTP Mix (10 mM ea.)
1 μ l	1 μ l	50X BD Advantage 2 Polymerase Mix
50 μl	50 μl	Total volume

3. Mix well and spin tube briefly to collect all the liquid in bottom of tube.
4. If your thermal cycler does not have a “hot lid”, add 1–2 drops of mineral oil to prevent evaporation during cycling. A good “seal” of mineral oil should have a well-defined meniscus between the two phases. Cap the PCR tubes firmly.
5. Commence thermal cycling using the following parameters:
 - 95°C for 1 min
 - 30 cycles*
 - 95°C for 15 sec
 - 68°C for 3 min

* 30 cycles with a 3-min annealing/extension time is sufficient for amplification of the positive control template provided in the kit. Other templates may require more or less cycles and different annealing/extension times (See Section V.B.).

6. Transfer a 5- μ l sample of your PCR reaction to a fresh tube and add 1 μ l of 5X stop/loading buffer. Analyze your sample(s), along with suitable DNA size markers, by electrophoresis on a 1.2% agarose/EtBr gel.

Expected results: If you are using the positive control reagents provided in the kit, the reaction should produce a single major fragment of 3.5 kb, derived from the gene for the bovine pancreatic trypsin inhibitor. No bands should be generated in the negative (i.e., no DNA template) control.

V. BD Advantage™ 2 PCR Procedure *continued*

B. Recommended Cycling Parameters for all BD Advantage™ 2 PCR Products

Use the following guidelines when setting up your initial experiments with the BD Advantage 2 system. These are general guidelines—the optimal parameters may vary with different thermal cyclers and will depend on your particular primers, template, and other experimental variables.

Notes:

- When using the BD Advantage 2 Polymerase Mix with our other PCR-based application kits (such as BD SMART cDNA Library Construction, BD SMART PCR cDNA Synthesis, BD Clontech PCR-Select Subtraction Kits, BD Marathon cDNA Amplification, Marathon-Ready cDNAs, or BD Delta Differential Display), use the parameters recommended in the protocol for that kit.
- If you intend to capture your PCR product by T/A cloning, we recommend that you add an additional 10-min extension at 70°C, and then immediately clone or freeze the PCR product. Do not store the reaction at 4°C. These steps will help ensure the incorporation and preservation of 3' A-overhangs.

Target Size	Cycle Parameters
<1 kb:	<ul style="list-style-type: none"> • 95°C for 1 min • 25–35 cycles^a <ul style="list-style-type: none"> 95°C for 30 sec^b 68°C for 1 min^c • 68°C for 1 min^d
1–5 kb:	<ul style="list-style-type: none"> • 95°C for 1 min • 25–35 cycles^a <ul style="list-style-type: none"> 95°C for 30 sec^b 68°C for 3 min^c • 68°C for 3 min^d
5–9 kb:	<ul style="list-style-type: none"> • 95°C for 1 min • 25–35 cycles^a <ul style="list-style-type: none"> 95°C for 30 sec^b 68°C for 6 min^c • 68°C for 6 min^d
10–20 kb:	<ul style="list-style-type: none"> • 95°C for 1 min • 25–35 cycles^a <ul style="list-style-type: none"> 95°C for 30 sec^b 68°C for 12 min^c • 68°C for 12 min^d

V. BD Advantage™ 2 PCR Procedure *continued*

- ^a 25 cycles for multiple-copy genes or medium-to-high abundance cDNAs; 30–35 cycles for single- or low-copy-number genes or rare cDNAs. For most applications, we prefer two-step cycles (denaturation at T_1 followed by annealing and extension at T_2) instead of three-step cycles (denaturation at T_1 followed by annealing at T_2 followed by extension at T_3). Three-step cycles will be necessary when the T_m of the primers is less than 60–65°C and in certain special protocols.
- ^b Use the minimal possible denaturation time. In some cases, better results may be obtained by modifying the denaturation step (94°C for 15 sec). Exposure of DNA to high temperatures causes some depurination of single-stranded DNA during denaturation, which eventually leads to strand scission. High temperature also leads to gradual loss of enzyme activity. Minimizing denaturation time is particularly important in experiments with very large templates where total cycling time can exceed 12 hr.
- ^c Use the maximum possible annealing/extension temperature. See “Note a”. Some researchers prefer to use an annealing/extension time equal to the expected target size (in kb) plus two minutes. We recommend using 1 min per kb of expected target.
- ^d Optional: This final extension may reduce background in some cases.

C. Recommendations for Electrophoresis

We recommend that you transfer a 5- μ l sample of your PCR reaction to a fresh tube and add 1 μ l of 5X stop/loading buffer. Place the remaining 45 μ l of the reaction mixture on ice; it can be subjected to further cycling if you do not see a product. Analyze your sample(s), along with suitable DNA size markers, by electrophoresis on a suitable agarose gel containing 0.1 μ g/ml EtBr. The percentage agarose and the DNA size markers you choose will depend on the expected range of insert sizes. You may wish to refer to the following general guidelines before assembling your gel.

Recommendations for agarose gels:

<u>Expected insert size range</u>	<u>Recommended % agarose</u>	<u>Recommended DNA size markers</u>
0.3 – 1.5 kb	1.5	ϕ X174/ <i>Hae</i> III
0.5 – 10 kb	1.2	1-kb DNA ladder
> 5 kb	0.8	λ / <i>Hind</i> III

D. Using 10X BD Advantage 2 SA PCR Buffer

In cases where non-specific background amplification is a problem, we strongly recommend the 10X BD Advantage 2 SA PCR Buffer be used as the buffer of choice.

Yields are unaffected when amplifying products less than 2 kb, whether the 10X BD Advantage 2 SA PCR Buffer or the BD Advantage 2 PCR Buffer is used. We recommend initial use of the 10X BD Advantage 2 PCR Buffer; and if background amplification arises, in particular with amplicons less than 2 kb in length, then we recommend switching to the 10X BD Advantage 2 SA PCR Buffer.

VI. Troubleshooting Guide

The following **general** guidelines apply to most PCR reactions. However, no attempt has been made to address troubleshooting for all of the many applications for which the BD Advantage 2 Polymerase Mix can be used. When using the enzyme with one of BD Biosciences Clontech companion products, additional application-specific troubleshooting information can be found in the relevant User Manual.

A. No product observed

PCR component missing or degraded	Use a checklist when assembling reactions. Always perform a positive control to ensure that each component is functional. If the positive control does not work, repeat the positive control only. If the positive control still does not work, repeat again replacing individual components to identify the faulty reagent.
Too few cycles	Increase the number of cycles (3–5 additional cycles at a time).
Annealing temp. too high	Decrease the annealing temperature in increments of 2–4°C.
Suboptimal primer design	Redesign your primer(s) after confirming the accuracy of the sequence information. If the original primer(s) was less than 22 nt long, try using a longer primer. If the original primer(s) had a G-C content of less than 45%, try to design a primer with a G-C content of 45–60%.
Not enough template	Repeat PCR using a higher concentration of DNA (after trying more cycles).
Poor template quality	Check template integrity by electrophoresis on a standard TBE-agarose gel. If necessary, repurify your template using methods that minimize shearing and nicking.
Denaturation temp. too high or low	Optimize denaturation temperature by decreasing or increasing it in 1°C increments. (A denaturation temperature that is too high can lead to degradation of the template, especially for long target sequences.)
Denaturation time too long or too short	Optimize denaturation time by decreasing or increasing it in 10-sec increments. (A denaturation time that is too long can lead to degradation of the template, especially for long target sequences.)
Extension time too short	Especially for longer templates, increase the extension time in 1-min increments.

VI. Troubleshooting Guide *continued*

Too little enzyme	The BD Advantage 2 Polymerase Mix is 50X for most applications; however a 1X final reaction concentration of the enzyme mix may be too low for some applications. Therefore, try to optimize the cycle parameters as described above before increasing the enzyme concentration. In rare cases, the yields can be improved by increasing the concentration of the enzyme mix. However, increasing the final concentration to >2X in the reaction mix is likely to lead to higher background levels.
[Mg ²⁺] is too low	The BD Advantage 2 Polymerase Mix performs well at a broad range of Mg ²⁺ concentration. Therefore, as long as you use the buffer included with the mix and a final concentration of 0.2 mM of each dNTP, it is unlikely that a lack of product is due to problems with the Mg ²⁺ concentration. However, high concentrations of EDTA or other metal chelators in the template stock solution can reduce the effective concentration of Mg ²⁺ to below a minimum level.
[dNTPs] is too low	<p>When used as recommended, the 50X dNTP mix provided with the kit gives a final concentration of 0.2 mM of each dNTP. In our experience, this concentration of dNTPs is suitable for a wide range of applications.</p> <p>If you are preparing your own dNTPs, be sure that your final concentration of each dNTP in the reaction is 0.2 mM.</p> <p>Some manufacturers recommend using concentrations higher than 0.2 mM of each dNTP when amplifying large templates. However, we have had no trouble amplifying large templates using 0.2 mM for each dNTP. Since we have gone up to 35 kb with the BD Advantage Genomic PCR Kit, it is unlikely that dNTPs are limiting. Note that if you increase the concentration of dNTPs, you will also need to increase the [Mg²⁺] proportionately.</p>
Difficult target	Some targets are inherently difficult to amplify. In most cases, this is due to unusually high G-C content and/or secondary structure. Use BD Advantage GC Polymerase Mix or Kit instead of the BD Advantage 2 system.

VI. Troubleshooting Guide *continued*

B. Multiple products

Too many cycles	Reducing the cycle number may eliminate nonspecific bands.
Annealing temp. too low	Increase the annealing/extension temperature in increments of 2–3°C.
Suboptimal primer design	Redesign your primer(s) after confirming the accuracy of the sequence information. If the original primer(s) was less than 22 nt long, try using a longer primer. If the original primer(s) had a G-C content of less than 45%, try to design a primer with a G-C content of 45–60%.
Touchdown PCR needed	“Touchdown” PCR significantly improves the specificity of many PCR reactions in various applications (Don <i>et al.</i> , 1991; Roux, 1995). Touchdown PCR involves using an annealing/extension temperature that is several degrees <i>higher</i> than the T_m of the primers during the initial PCR cycles. The annealing/extension temperature is then reduced to the primer T_m for the remaining PCR cycles. The change can be performed either in a single step or in increments over several cycles.
Contamination	See Section E below.

C. Low yield

Poor template quality	Check template integrity by electrophoresis on a standard TBE-agarose gel. If necessary, repurify your template using methods that minimize shearing and nicking.
-----------------------	---

D. Products are smeared on gel

Too many cycles	Reduce the cycle number by 3–5 to see if non-specific bands go away.
Denaturation temp. too low	Try increasing the denaturation temperature in increments of 1°C.
Extension time too long	Decrease the extension time in 1–2-min increments.
Poor template quality	Check template integrity by electrophoresis on a de-denaturing agarose gel. Repurify your template if necessary.
Touchdown PCR needed	See “Touchdown PCR needed” under previous section.

VI. Troubleshooting Guide *continued*

Non-specific priming	Use the 10X BD Advantage SA PCR Buffer which has been specifically engineered to address this (rather than the 10X BD Advantage 2 PCR Buffer).
Too much enzyme	The BD Advantage 2 Polymerase Mix is 50X for most applications; however, a 1X final concentration of the enzyme mix may be too high for some applications. If smearing is observed, first try optimizing the cycle parameters as described above, then try reducing the enzyme concentration to 0.5–0.2X.
Too much template	Try a lower concentration of DNA template in the PCR reaction.
Contamination	See Section E below.

E. Dealing with contamination

Contamination most often results in extra bands or smearing. It is important to include a negative control (a control that replaces the DNA template with PCR-grade H₂O but still includes the primers) in every PCR experiment to determine if the PCR reagents, pipettors or PCR reaction tubes are contaminated with previously amplified targets.

If possible, set up the PCR reaction and perform the post-PCR analysis in separate laboratory areas with separate sets of pipettors.

Laboratory benches and pipettor shafts can be decontaminated by depurination. Wipe surfaces with 1N HCl followed by 1N NaOH. Then neutralize with a neutral buffer (e.g., Tris or PBS) and rinse with ddH₂O.

We advise using commercially available aerosol-free pipette tips.

An enzymatic method has been published for destroying PCR product carryover (Longo *et al.*, 1990). It involves incorporation of dUTP into the PCR products and subsequent hydrolysis with uracil-N-glycosylase (UNG).

When performing PCR directly on phage plaques or bacterial colonies, failure to isolate single plaques or colonies will also produce multiple bands.

VII. References

- Barnes, W. M. (1994) PCR amplification of up to 35-kb DNA with high fidelity and high yield from λ bacteriophage templates. *Proc. Natl. Acad. Sci. USA* **91**:2216–2220.
- Cheng, S., Fockler, C., Barnes, W. M. & Higuchi, R. (1994) Effective amplification of long targets from cloned inserts and human genomic DNA. *Proc. Natl. Acad. Sci. USA* **91**:5695–5699.
- Don, R. H., Cox, P. T., Wainwright, B. J., Baker, K. & Mattick, J. S. (1991) 'Touchdown' PCR to circumvent spurious priming during gene amplification. *Nucleic Acids Res.* **19**:4008.
- Frey, B. & Suppmann, B. (1995) Demonstration of the Expand™ PCR system's greater fidelity and higher yields with a *lacI*-based PCR fidelity assay. *Biochemica* **2**:8–9.
- Kellogg, D. E., Rybalkin, I., Chen, S., Mukhamedova, N., Vlasik, T., Chenchik, A. & Siebert, P. (1994) TaqStart: A Neutralizing Monoclonal Antibody that facilitates "Hot Start" PCR. *Clontechiques IX*(2):1–5.
- Kellogg, D. E., Rybalkin, I., Chen, S., Mukhamedova, N., Vlasik, T., Siebert, P. & Chenchik, A. (1994) TaqStart Antibody: Hotstart PCR facilitated by a neutralizing monoclonal antibody directed against *Taq* DNA polymerase. *BioTechniques* **16**:1134–1137.
- Longo, M. C., Berninger, M. S. & Hartley, J. L. (1990) Use of uracil DNA glycosylase to control carry-over contamination in polymerase chain reactions. *Gene* **93**:3749.
- Nelson, K., Brannan, J. & Kretz, K., (1995) The fidelity of TaqPlus™ DNA Polymerase in PCR. *Strategies Mol. Biol.* **8**:24–25.
- Roux, K. H. (1995) Optimization and troubleshooting in PCR. *PCR Methods Appl.* **4**:5185–5194.
- Sambrook, J., Fritsch, E. F. & Maniatis, T. (1989) *Molecular Cloning: A Laboratory Manual, Second Edition* (Cold Spring Harbor Laboratory, Cold Spring Harbor, NY).

VIII. Related Products

For a complete listing of all BD Biosciences Clontech products, please visit www.bdbiosciences.com/clontech

PCR Products	Cat. No.	New Cat. No.
• BD Sprint™ TITANIUM™ <i>Taq</i> 384 Plate	K1952-1	639552
• BD Sprint™ Advantage™ 96 Plate	K1950-1	639550
• BD Sprint™ Advantage™ Single Shots	K1953-y	639556
	K1953-1	639553
	K1953-2	639554
• 10X BD Advantage 2 PCR Buffer	8207-1	639137
	8207-2	639138
• BD TITANIUM™ <i>Taq</i> DNA Polymerase	8434-1	639208
	8434-2	639209
• BD TITANIUM™ <i>Taq</i> PCR Kit	K1915-y	639211
	K1915-1	639210
• BD Advantage™ Genomic Polymerase Mix	8418-1	639110

VIII. Related Products *continued*

PCR Products <i>continued</i>	Cat. No.	New Cat. No.
• BD Advantage™ Genomic PCR Kit	K1906-y	639104
	K1906-1	639103
• BD Advantage™ GC 2 Polymerase Mix	8433-1	639114
• BD Advantage™ GC 2 PCR Kit	K1913-y	639120
	K1913-1	639119
• BD Advantage™ GC Genomic Polymerase Mix	8420-1	639113
• BD Advantage™ GC Genomic PCR Kit	K1908-y	639118
	K1908-1	639117
• BD Advantage™ HF 2 PCR Kit	K1914-y	639124
	K1914-1	639123
• BD Advantage™ UltraPure PCR Deoxynucleotide Mix	4700-1	639125
• BD Advantage™ UltraPure dNTP Combination Kit	4715-1	639132
• BD TaqStart™ Antibody	5400-1	639250
	5400-2	639251
PCR Amplification Kits & Products		
• BD SMART™ cDNA Library Construction Kit	K1051-1	634901
• BD Super SMART™ PCR cDNA Synthesis Kit	K1054-1	635000
• BD SMART™ RACE cDNA Amplification Kit	K1811-1	634914
• BD Marathon™ cDNA Amplification Kit	K1802-1	634913
• BD™ Marathon-Ready cDNAs	many	many
• BD Delta™ Differential Display Kit	K1810-1	637405
• BD Clontech PCR-Select™ cDNA Subtraction Kit	K1804-1	637401
• BD Clontech PCR-Select™ Bacterial Genome Subtraction Kit	K1809-1	637404
• LD-Insert Screening Amplimer Sets	many	many
• RT-PCR Amplimer Sets	many	many
• BD Advantage™ RT-for-PCR Kit	K1402-1	639505
	K1402-2	639506
• BD MTC™ Multiple Tissue cDNA Panels	many	many
• BD QUICK-Clone™ cDNAs	many	many
• BD TITANIUM™ One-Step RT-PCR Kit	K1403-1	639503
	K1403-2	639504

Notes

Notes

Notes

Notice to Purchaser

This product is intended to be used for research purposes only. It is not to be used for drug or diagnostic purposes nor is it intended for human use. BD Biosciences Clontech products may not be resold, modified for resale, or used to manufacture commercial products without written approval of BD Biosciences Clontech.

BD Advantage™ 2 products are covered by U.S. Patent #5,436,149.

BD Biosciences Clontech PCR-Select™ cDNA Subtraction products are covered by U.S. Patent Nos. 5,565,340 and 5,759,822, as well as pending foreign patent applications. For-profit and not-for-profit purchasers of PCR-Select products are entitled to use the reagents for research, including identification of molecular markers or differentially expressed genes; however, the following uses expressly are prohibited: (1) performing services for third parties; (2) identifying nucleic acid sequences to be included on nucleic acid arrays, blots or in libraries or other cDNA collections which are then sold to third parties; or (3) constructing databases which are then sold to third parties. Further, reproduction, amplification, modification, reformulation or resale of the reagents provided in the PCR-Select products is not permitted. For information on licensing PCR-Select for these purposes, please contact the Product Manager for PCR Applications by mail at BD Biosciences Clontech, 1020 East Meadow Circle, Palo Alto, CA, 94303; or by phone at 650-424-8222 (extension 7816); or by e-mail at licensing@clontech.com.

BD SMART™ technology is covered by U.S. Patents #5,962,271 & 5,962,272.

BD TaqStart™ Antibody is licensed under U.S. Patent #5,338,671 and corresponding patents in other countries.

A license under U.S. patents 4,683,202, 4,683,195 and 4,965,188 or their foreign counterparts owned by Roche Molecular Systems, Inc. and F. Hoffmann-La Roche Ltd. ("Roche"), for use in research and development, has an up-front fee component and a running-royalty component. By purchasing certain Applied Biosystems PCR reagents, purchasers obtain limited, nontransferable rights under the running royalty component to use the purchased amount of reagents to practice the Polymerase Chain Reaction (PCR) and related processes described in said patents only for research and development activities of the purchaser. Such use is fully licensed under these patents only when the reagents are used in conjunction with a thermal cycler whose use is covered by the up-front fee component. Rights under the up-front fee component of a PCR license may be purchased from Applied Biosystems or obtained by purchasing an Authorized Thermal Cycler. No right to perform or offer commercial services of any kind using PCR, including, for example, reporting the results of purchaser's activities for a fee or other commercial consideration, is granted unless the purchaser obtains a further license from either Applied Biosystems or Roche. Additional information on purchasing licenses to practice the PCR process may be obtained by contacting the Director of Licensing at Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404 or Roche Molecular Systems, 1145 Atlantic Avenue, Alameda, California, 94501.

GeneAmp® is a registered trademark of Roche Molecular Systems, Inc., licensed to the Perkin-Elmer Corporation.

BD, BD Logo and all other trademarks are property of Becton, Dickinson and Company. ©2003 BD

